



Thrombodynamics Analyser

Fibrinolysis assay

Patient:		04.11.2019 11:18
Age:	, History N:	
Measurement title:	1	MIX
Plasma type:	Frozen PFP	
Sample code:	HIFIVE - HF25 V1	
Physician:	, Medical Institution/Department:	
Diagnosis:	, Pregnancy, weeks:	
Prescribed medications:		

Fibrin dynamics parameters	Unit	Parameter	Value	Norm.
Rate of clot growth	um/min	V	32.6	
Lag time	min	Tlag	1.0	
Initial rate of clot growth	um/min	Vi	64.8	
Stationary rate of clot growth	um/min	Vst		
Clot size at 30 min	um	CS	1278	
Clot density	a.u.	D	21649	
Spontaneous clotting	min	Tsp		

Fibrinolysis parameters	Unit	Parameter	Value	Norm.
Lysis onset time	min	LOT	20.0	
Lysis progression	%/min	LP	21.1	



Comment:

Tech info:

ATTENTION: Research use only. Measurement results shouldn't be used for diagnostic purposes.

S/N: 250018, Operator: Admin, report created 11.11.2019 15:02

