



# Thrombodynamics Analyser

## Fibrinolysis assay

<b>Patient:</b>		<b>30.10.2019 09:37</b>
<b>Age:</b>	, History N:	
Measurement title:	3	<b>MIX</b>
Plasma type:	Frozen PFP	
Sample code:	HIFIVE - HF22 V3	
Physician:	, Medical Institution/Department:	
Diagnosis:	, Pregnancy, weeks:	
Prescribed medications:		

Fibrin dynamics parameters	Unit	Parameter	Value	Norm.
Rate of clot growth	um/min	V	37.8	
Lag time	min	Tlag	1.1	
Initial rate of clot growth	um/min	Vi	60.5	
Stationary rate of clot growth	um/min	Vst	37.8	
Clot size at 30 min	um	CS	1370	
Clot density	a.u.	D	20590	
Spontaneous clotting	min	Tsp	46.1	

Fibrinolysis parameters	Unit	Parameter	Value	Norm.
Lysis onset time	min	LOT	36.5	
Lysis progression	%/min	LP	10.9	



Comment:
----------

Tech info:
------------

