



# Thrombodynamics Analyser

## Fibrinolysis assay

<b>Patient:</b>		<b>30.10.2019 07:50</b>
<b>Age:</b>	, History N:	
Measurement title:	1	<b>MIX</b>
Plasma type:	Frozen PFP	
Sample code:	HIFIVE - HF22 V1	
Physician:	, Medical Institution/Department:	
Diagnosis:	, Pregnancy, weeks:	
Prescribed medications:		

Fibrin dynamics parameters	Unit	Parameter	Value	Norm.
Rate of clot growth	um/min	V	39.5	
Lag time	min	Tlag	0.9	
Initial rate of clot growth	um/min	Vi	62.9	
Stationary rate of clot growth	um/min	Vst	39.5	
Clot size at 30 min	um	CS	1479	
Clot density	a.u.	D	22013	
Spontaneous clotting	min	Tsp	25.3	

Fibrinolysis parameters	Unit	Parameter	Value	Norm.
Lysis onset time	min	LOT	34.1	
Lysis progression	%/min	LP	12.4	



Comment:
Tech info: