



# Thrombodynamics Analyser

## Fibrinolysis assay

<b>Patient:</b>		<b>07.11.2019 12:47</b>
<b>Age:</b>	, History N:	
Measurement title:	3	<b>MIX</b>
Plasma type:	Frozen PFP	
Sample code:	HIFIVE - HF31 V3	
Physician:	, Medical Institution/Department:	
Diagnosis:	, Pregnancy, weeks:	
Prescribed medications:		

Fibrin dynamics parameters	Unit	Parameter	Value	Norm.
Rate of clot growth	um/min	V	27.9	
Lag time	min	Tlag	0.7	
Initial rate of clot growth	um/min	Vi	60.6	
Stationary rate of clot growth	um/min	Vst	27.9	
Clot size at 30 min	um	CS	1221	
Clot density	a.u.	D	28189	
Spontaneous clotting	min	Tsp		

Fibrinolysis parameters	Unit	Parameter	Value	Norm.
Lysis onset time	min	LOT	23.5	
Lysis progression	%/min	LP	15.7	



Comment:
----------

Tech info:
------------

ATTENTION: Research use only. Measurement results shouldn't be used for diagnostic purposes.

S/N: 250018, Operator: Admin, report created 11.11.2019 15:18

